

Index

abscess case study, 146–7	accessibility issues	mobile mammography, 86	workplace mammography, 86
active matrix liquid crystal display (AMLCD)	technology, 28	alternative forced choice (AFC)	methodology, 4–5
ambient light, 59	ambient temperature, mobile units, 91–2	amorphous selenium (a-Se)	detectors, 8, 20
amorphous silicon (a-Si)	detectors, 7, 20	application service provider (ASP), 36–7	architectural distortion, 110, 117
case studies, 155–6, 166–7	archiving, 37. <i>See also</i> picture archiving and communication systems (PACS)	case study, 41–5	data compression, 37–8
influencing factors, 37	growth and new technology, 37	image size, 37	processed vs. raw images, 37
system evaluation and assessment, 39–41	availability, 39	data migration, 41	data security, 40
fault tolerance, 39	integration, 40	performance, 41	scalability, 39
system management, 40–1	artifacts, 71	breast CT, 130	detector-related, 71–5
banding artifacts, 72–4	ghosting artifacts, 74–5	misread lines, 72	multiple dead detector elements, 72
single dead detector elements, 71–2	gantry-related, 75	foreign material on the compression paddle, 75–6	grid artifacts, 76
improper collimation, 77–8	patient-related, 78–81	foreign substances on skin, 80–1	patient motion, 78–9
unwanted anatomy, 79	processing-related, 81–4	gain and offset calibration artifacts, 83–4	pixel reversal artifacts, 82–3
shading artifacts, 82	skin line processing algorithm failure, 81	storage-related, 84	atypical ductal hyperplasia (ADH)
stereotactic biopsy, 96	Audit Trail and Node Authentication (ATNA), 40	authentication, 40	authorization, 40
backing up data, 37	offsite backups, 36–7	banding artifacts, 72–4	bandwidth, 34
biopsy. <i>See</i> stereotactic biopsy	bit, 18	bloom artifact case study, 186	Boise mobile mammography case study, Idaho, 88–94
computed radiography (CR) digital mammography advantages and disadvantages, 91	conversion to FFD unit, 90–1	cost analysis, 94	film-screen unit initiation, 90
initial quality assurance, 90	operational details, 93–4	problems with digital machines in mobile units, 91–3	breast abscess case study, 146–7
breast cancer mortality trends, 85	breast compression, 112	digital breast tomosynthesis, 112–13	breast computed tomography (breast CT)
background, 125	contrast use, 140–1	correlation with other modalities, 133–40	mammography, 133–40
PET/CT, 140	tomosynthesis, 140	detector, 127–8	future improvements, 141–2
dose reduction, 142	improved resolution, 142	gantry and table, 127	general design, 125–7
image quality, 128–30	contrast, 128	noise level, 129–30	signal accuracy and artifacts, 130
spatial resolution, 128–9	microcalcification detection, 125, 128, 131–6	radiation therapy planning, 141	tissue structures, 130–1
x-ray source, 127	x-ray techniques and doses, 133	byte, 18	calcifications, 134, 136
bloom artifact, 186	case studies, 148–50, 154–7, 159, 162, 175, 177, 185, 186	digital breast tomosynthesis and, 120–1	fibroadenoma, 159, 185
fibrocystic, 154–5	linear, 175	microcalcification detection, 68	breast CT, 125, 128, 131–6
specimen radiography, 104	needle localization, 100	pleomorphic, 98, 110, 137	pseudocalcifications, 162–3
residual capsule following implant removal, 175	secretory, 156–7	skin, 177	stereotactic biopsy, 99
callback rate, 64	digital breast tomosynthesis advantages, 115	carcinoma. <i>See also</i> ductal carcinoma in situ (DCIS)	invasive, 68, 132
case studies, 145	ductal, 106, 118, 133, 138, 167	lobular, 171	metastatic, 135
mucinous, 105	cathode ray tube (CRT) monitors, 24	charge-coupled devices (CCDs), 8	clip markers, 99
collimation errors, 77–8	communication between system components, 32–4	complementary metal-oxide semiconductors (CMOS), 8	compliance problems, 86
possible solutions, 86	addressing convenience and accessibility, 86	cost reduction, 86	mobile mammography, 86
compression. <i>See</i> breast compression; data compression			

- computed radiography (CR), 9–12, 20
  - advantages and disadvantages, 91
- computed tomography (CT). *See* breast computed tomography
- computer-aided detection (CADE), 1, 53–9, 92–3
  - commercially available systems, 187
- digital breast tomosynthesis application, 123
- time taken, 58
- usage of, 52
- variation between systems, 57
- computer languages, 32–4
- cone-beam computed tomography (CBCT), 125
- confidentiality, 40
- contrast detail phantom, 4
- contrast-limited adaptive histogram equalization (CLAHE), 23–4
- cost analysis, mobile units, 94
- cupping artifacts, 128, 130
- cut-off spatial frequency, 2
- cyst, 116
  - aspiration case study, 163
- data compression, 37–8
- 
- 
- 
- lossless compression, 38
- lossy compression, 38
- data migration, 41
- data redundancy, 40
- data security, 39, 40
  - authentication, 40
  - authorization, 40
  - confidentiality, 40
  - encryption, 40
  - integrity, 40
- data transmission, 38–9
  - telecommunications carrier, 39
- deodorant, 80
- detective quantum efficiency (DQE), 4, 129
- detector elements (DEs), 20
  - multiple dead detector elements, 72
  - single dead detector elements, 71–2
- detectors, 20. *See also* artifacts; detector elements (DEs)
- breast CT, 127–8
- development of, 1–2
- digital cassette
  - mammography, 9–12
- direct conversion, 8–9, 20
- indirect conversion, 20
  - with prompt readout, 6–8
- mobile units, 91–2
- photon counting mammography, 12
- developing mass case study, 160–1
- digital breast tomosynthesis, 110–11
  - breast compression, 112–13
  - breast CT correlation, 140
  - clinical implementation, 122
  - commercially available systems, 187
  - contrast-enhanced, 123
  - dose, 113
  - equipment, 114–15, 122
  - FDA approval, 115, 122
  - future applications, 122–3
  - image acquisition, 111–13
  - image interpretation, 113–14
  - time taken, 114
  - image reconstruction, 113
  - literature review, 118–22
  - calcifications, 120–1
  - diagnostic performance, 119–20
  - masses and asymmetries, 121–2
  - reader preference, 119
  - potential advantages, 115–17
  - decreased recall rate, 115
  - localization of lesion, 115
  - mass characterization, 115
  - mass detection, 115
  - potential disadvantages, 117–18
  - costs, 117
  - interpretation time, 117
  - lack of uniformity, 117
  - not all breasts fit the detector, 118
  - training requirements, 117
  - workflow issues, 122
- digital cassette
  - mammography, 9–12
- Digital Imaging and Communications in Medicine (DICOM), 32–4
- gray-scale display function (GSDF), 30
- standards, 29
- Digital Mammographic Imaging Screening Trial (DMIST), 22, 63
- digital mammography, 18, 109. *See also* interpretation of digital mammograms; mammography; mobile mammography
- available units, 19, 187
- development of, 1–2, 85
- detectors, 1–2
- efficacy, 63–9
- future trends, 13–16
- physics of, 18
- upgrading a film-screen mobile unit to FFDM, 88
- versus film-screen mammography, 20–1, 88, 102–4
- practical improvements, 104
- speed improvement, 104
- digital radiography (DR), 20
- digital telemammography, 88
- digital visual interface (DVI), 28
- digitization of film
  - mammograms, 1, 51
- direct access storage (DAS), 35
- direct conversion detectors, 8–9, 20
- double reading, 53–9
- dual-side read CR
  - mammography, 12
- ductal carcinoma in situ (DCIS), 141
  - case studies, 148–9, 152–4, 167, 179, 181–4
  - detection rate, 68
  - stereotactic biopsy, 96
- ductography, 99
- edge spread function (ESF), 2
- electronic health records (EHRs), 39
- electronic personal health information (ePHI), 39
- encryption, 40
- enterprise-wide image distribution, 104
- fault tolerance, 39
- fibroadenoma
  - calcified, 159, 185
  - case studies, 151–2, 185
- film-screen mammography, 1, 18, 109
  - efficacy, 62
  - image display, 27
  - upgrading a mobile unit to FFDM, 88
  - vs. digital mammography, 20–1
- focal spot blurring effect, 129
- full-field digital mammography (FFDM), 8. *See also* digital mammography
- conversion to, case study, 90–1
- image size, 37
- upgrading a film-screen mobile unit to, 88
- gain and offset calibration artifacts, 83–4
- galactography, 99
  - improvements with digital mammography, 106
- geometric magnification, 129
- ghosting artifacts, 74–5
- granulomatous mastitis case study, 165
- graphics cards, 27–30
- grid artifacts, 76
- hamartoma case study, 173
- hanging protocols, 25–6, 50–1
- hard-copy display, 25
- hardware, 34–5
  - network infrastructure, 34–5
  - storage infrastructure, 35
- Health Insurance Portability and Accountability Act (HIPAA), 39
- Health Level 7 (HL7), 32
- hematoma, 105, 107, 136
- image acquisition, 18–20. *See also* detectors
- digital breast tomosynthesis (DBT), 111–13
- image display, 24–5, 27
  - enterprise-wide distribution, 104
  - hanging protocols, 25–6, 50–1
  - monitors, 28–9
  - classification, 29
  - image size considerations, 47–50
  - quality management, 29–31
  - workstations, 27–8
- image interpretation. *See* interpretation of digital mammograms
- image management, 41
- image plate (IP), 9
- image processing, 22–4, 51–3, 104
  - contrast-limited adaptive histogram equalization, 23–4
  - intensity windowing, 22–3, 52–3
  - peripheral equalization, 23
  - processing-related artifacts. *See* artifacts
  - unsharp masking, 23
- image quality
  - breast CT, 128–30
  - metrics, 2–6
- image size, 37, 47–50
- image transmission, 88
- mobile units, 92
- implant removal
  - calcified residual capsule, 175
  - silicone mass, 158–9
- indirect conversion detectors, 20
  - with prompt readout, 6–8
  - photodetector, 7–8
  - scintillator, 6
- inflammatory mass case study, 146–7
- Integrating the Healthcare Enterprise (IHE) initiative, 32, 40
- intensity windowing, 22–3, 52–3
- interobserver variability, 68

## Index

- interpretation of digital  
 mammograms, 46  
 computer-aided detection  
 (CADE), 53–9  
 digital breast tomosynthesis,  
 113–14  
 double reading, 53–9  
 hanging protocols, 50–1  
 image size choice, 47–50  
 post-processing at the  
 monitor, 51–3  
 reading environment, 59  
 time utilized, 46–7  
 intraobserver variability, 68  
 invasive cancer detection, 68  
 case studies, 145, 167, 169,  
 171, 179  
 invasive ductal carcinoma,  
 167, 169, 179  
 invasive lobular carcinoma, 171
- latitude, 18  
 line conditioner, 92  
 line spread function (LSF), 2  
 liquid crystal display (LCD)  
 monitors, 24, 27–9  
 look-up tables (LUTs), 28
- mammographically guided  
 procedures, 96  
 future directions, 107  
 galactography, 99  
 improvements with digital  
 mammography, 104–6  
 needle localization, 99  
 specimen radiography, 101–2  
 stereotactic biopsy, 96–9  
 mammography, 109. *See also*  
 digital mammography;  
 film-screen  
 mammography; mobile  
 mammography  
 breast CT correlation, 133–40  
 compliance problems, 86  
 possible solutions, 86  
 efficacy, 62  
 limitations, 110, 133  
 technology evolution, 85  
 timing, 62–3
- Medicare, 63, 87  
 metal clip artifacts, 130  
 microcalcification detection, 68  
 breast CT, 125, 128, 131–6  
 case study, 148–9  
 specimen radiography, 104  
 misread lines, 72  
 mobile mammography.  
*See also* Boise mobile  
 mammography case  
 study, Idaho  
 compliance benefits, 86  
 digital tele mammography, 88  
 portable mammography  
 units, 86
- true mobile units, 86–8  
 upgrading a film-screen  
 mobile unit to FFDM, 88  
 modulation transfer function  
 (MTF), 2, 20  
 monitors, 28–9, 92. *See also*  
 liquid crystal display  
 (LCD) monitors  
 classification, 29  
 image size considerations,  
 47–50  
 quality management, 29–31  
 frequency of QC checks, 31  
 mortality trends, 85
- needle localization, 99  
 improvements with digital  
 mammography, 105  
 network attached storage  
 (NAS), 35  
 network infrastructure, 34–5  
 noise equivalent quanta (NEQ), 4  
 noise level, breast CT, 129–30  
 noise power spectrum (NPS), 3  
 normalized noise power  
 spectrum (NPS), 3  
 Nyquist frequency, 2
- oil cyst case study, 147–8  
 optical magnification use, 53  
 organic light-emitting diode  
 (OLED), 27
- papilloma, 102, 106  
 parallax shift, 111, 112  
 patient compliance.  
*See* compliance problems  
 pendant geometry, 125, 127  
 peripheral equalization, 23  
 personal health information  
 (PHI), 39  
 photomultiplier tube (PMT), 12  
 photon-counting  
 mammography, 12  
 picture archiving and  
 communication  
 systems (PACS), 27.  
*See also* archiving  
 monitors, 29  
 pixel, 18  
 pixel binning, 15  
 pixel pitch, 2  
 pixel reversal artifacts, 82–3  
 pneumocystography case  
 study, 163  
 point spread function (PSF), 2  
 portable mammography  
 units, 86  
 positive predictive value, 64–8  
 post-processing. *See* image  
 processing  
 power supply, mobile units, 92  
 processed images, 37  
 project management, 41
- pseudoangiomatous stromal  
 hyperplasia (PASH), 139  
 case study, 151–2  
 pseudocalcifications, 162–3
- quality assurance (QA)  
 Boise mobile mammography  
 case study, Idaho, 90  
 monitors, 29–31  
 quality control (QC)  
 monitors, 29–31  
 frequency of QC  
 checks, 31
- radial scar case study, 150–1  
 radiation therapy planning,  
 141  
 raw images, 37  
 reading environment, 59  
 ambient light, 59  
 rebinning, 47  
 recall rate, 64  
 digital breast tomosynthesis  
 advantages, 115  
 receiver operating characteristic  
 (ROC), 5  
 redundant array of independent  
 disks (RAID), 35  
 replication, 40
- scalability of archiving  
 system, 39  
 scheduling, mobile units, 93  
 screen-film mammography.  
*See* film-screen  
 mammography  
 screening mammography.  
*See* digital  
 mammography; film-  
 screen mammography;  
 mammography
- seroma, 134  
 shading artifacts, 82  
 signal difference to noise ratio  
 (SDNR), 4  
 silicone mass from ruptured  
 implant, 158–9
- skin  
 calcifications, 177  
 foreign substances  
 on, 80–1  
 visualization of, 104  
 skin-line processing algorithm  
 failure, 81  
 skin mole case study, 164  
 snowstorm appearance,  
 158–9  
 spatial resolution, 2, 19  
 breast CT, 128–9  
 specimen radiography, 101–2,  
 110  
 improvements with digital  
 mammography, 105–6  
 staffing, mobile units, 93
- step-and-shoot image acquisition  
 method, 111  
 stereotactic biopsy, 96–9  
 improvements with digital  
 mammography, 104–5
- storage  
 data compression, 37–8  
 influencing factors, 37  
 growth and new  
 technology, 37  
 image size, 37  
 processed vs. raw images, 37  
 infrastructure, 35  
 direct access storage  
 (DAS), 35  
 network attached storage  
 (NAS), 35  
 removable media, 35  
 storage area network  
 (SAN), 35  
 redundant array of independent  
 disks (RAID), 35  
 storage-related artifacts, 84  
 tiers, 35–7  
 offsite backups, 36–7  
 offline or long-term  
 storage, 36  
 online or short-term  
 storage, 35–6  
 storage area network (SAN), 35  
 online or short-term storage, 36  
 system maintenance, 41
- telecommunications carrier, 39  
 tele mammography, 88  
 digital breast tomosynthesis  
 application, 123  
 thin-film transistor (TFT), 9, 20  
 timing of screening  
 mammography, 62–3  
 tomography, 111. *See also* breast  
 computed tomography  
 tomosynthesis. *See also* digital  
 breast tomosynthesis  
 Trex processing, 23  
 tubular carcinoma case study, 166
- underexposure, 157–8  
 uninterruptible power supply  
 (UPS), 92  
 unsharp masking, 23
- vacuum-assisted device (VAD), 98  
 video graphics array (VGA), 28  
 virtual private network (VPN), 40
- wire localization, 99  
 workplace mammography, 86  
 workstations, 27–8
- Xerox mammography, 85
- zooming, 51–3. *See also* image  
 size